

CERTIFICATE OF CALIBRATION

Unit Model #: 195374

Unit Serial #: LG030068G

Recorder Serial #: UR34925-001-003 0306 K28

Instrument #: 29686

MAC Address: 08:00:48:80:73:F6

Sure Controls does hereby certify the above instrument was calibrated against standards maintained by Sure Controls and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Displayed Value

Recorder	Primary Standard Signal 382.0°F	Primary Standard Signal 882.0°F	Primary Standard Signal 1382.0°F
CHANNEL 1	382	883	1384
CHANNEL 2	382	883	1383
CHANNEL 3	383	883	1384
CHANNEL 4	382	882	1383
CHANNEL 5	382	883	1384
CHANNEL 6	382	883	1384

Calibration Date: 4/4/2011

Recalibration Date: 4/4/2012

Certified by: JAB

Instrumentation Used: Beamex Model No. MC5 Multi-Function Calibrator

Beamex Serial No. 25513615MJP

Sure Controls ID No. STD-035