

## **CERTIFICATE OF CALIBRATION**

Unit Model #:	195374
Unit Serial #:	LH050051G
Recorder Serial #:	UR37481-1-7-0507-B5
Instrument #:	52001
MAC Address:	00:0A:8D:00:CB:21

Sure Controls does hereby certify the above instrument was calibrated against standards maintained by Sure Controls and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

	Displayed Value		
Recorder	Primary Standard Signal 382.0 °F	Primary Standard Signal 882.0°F	Primary Standard Signal 1382.0°F
CHANNEL 1	381	882	1382
CHANNEL 2	382	883	1384
CHANNEL 3	381	881	1382
CHANNEL 4	382	883	1384
CHANNEL 5	382	882	1383
CHANNEL 6	381	881	1382
Calibration Date: Recalibration Date: Certified by:	3/22/2012 3/22/2013 JAB		

Instrumentation Used: Beamex Model No. MC5 Multi-Function Calibrator Beamex Serial No. 25513615MJP Sure Controls ID No. STD-035