

Red-D-Arc Welderrentals

Calibration Certificate

Unit Serial Number: LH390014G

Date Tested: 2/2/2015

Instrument Number: 49740

Mac Address: 00:0A:8D:00:C2:4C

IP Address: 192.168.111.222

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	2.773	379	6.009	881	9.230	1381
Channel 2	2.782	381	6.009	881	9.228	1381
Channel 3	2.781	380	6.009	881	9.232	1381
Channel 4	2.780	380	6.012	881	9.235	1381
Channel 5	2.780	380	6.016	882	9.237	1382
Channel 6	2.780	380	6.016	882	9.237	1382

Allowable difference in temp reading is +or- 5 Deg. F.

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 1882015

Calibration Date: 10/21/14

Precision digital multimeter SN: 2036021

Calibration Date: 10/21/14

ProHeat 35 power supply used for signal output SN: N/A

Calibration date: N/A

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Technician: Scott T.

Calibration due date: 2/2/2016

Signature