

Red-D-Arc Welderrentals

Calibration Certificate

Unit Serial Number: LJ040058G

Date Tested: 6/21/13

Instrument Number: 51380

Mac Address: 00:0A:8D:00:C8:B4

IP Address: 192.168.111.222

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	2.786	382	6.008	882	9.235	1382
Channel 2	2.782	382	6.010	882	9.236	1382
Channel 3	2.783	382	6.010	882	9.231	1382
Channel 4	2.788	382	6.008	882	9.236	1382
Channel 5	2.786	382	6.009	882	9.232	1382
Channel 6	2.788	382	6.009	882	9.233	1382

Allowable difference in temp reading is +or- 5 Deg. F.

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2083066

Calibration Date: 6/25/12

Precision digital multimeter SN: 2141018

Calibration Date: 8/15/12

ProHeat 35 power supply used for signal output SN: MA370104G

Calibration date: 6/21/13

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Technician: Don Wagner

Calibration due date: 6/21/14

Signature