Red-D-Arc Welderentals

Calibration Certificate

Unit Serial Number: LJ380146G Date Tested: 02-02-17

Instrument Number: 63739 Mac Address: 00:0A:8D:00:F8:FB

IP Address: 192.168.111.222

Primary Standard							
	382 (F)		882 (F)		1382 (F)		
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display	
Channel 1	2.791	383	6.016	882	9.240	1382	
Channel 2	2.796	383	6.019	883	9.244	1382	
Channel 3	2.796	383	6.023	883	9.247	1382	
Channel 4	2.796	383	6.020	883	9.246	1382	
Channel 5	2.797	383	6.020	883	9.244	1382	
Channel 6	2.796	383	6.027	884	9.252	1383	

Allowable difference in temp reading is +or- 5 Deg. F.

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 1897018

Calibration Date: 11-04-16

Precision digital multimeter SN: US36049893

Calibration Date: 11-04-16

ProHeat 35 power supply used for signal output SN: MC100331G

Calibration date: 02-01-17

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.	X	Technician:	Jared Lamar
Calibration due date:	02-02-18		Signature