

## CERTIFICATE OF CALIBRATION

**Unit Model #:** 195374

Unit Serial #: LJ380151G

Recorder Serial #: US25811-1-7-1009-PL1

Instrument #: 63737

**MAC Address:** 00:0A:8D:00:F8:F9

Sure Controls does hereby certify the above instrument was calibrated against standards maintained by Sure Controls and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

## **Displayed Value**

Recorder	Primary Standard Signal 382.0°F	Primary Standard Signal 882.0°F	Primary Standard Signal 1382.0°F
CHANNEL 1	382	882	1382
CHANNEL 2	382	882	1382
CHANNEL 3	382	882	1382
CHANNEL 4	382	882	1382
CHANNEL 5	382	882	1382
CHANNEL 6	382	882	1382

Calibration Date: 9/2/2009

Recalibration Date: 9/2/2010

Certified by: VH

Instrumentation Used: Beamex Model No. MC5 Multi-Function Calibrator

Beamex Serial No. 25513615MJP

Sure Controls ID No. STD-035