## **Red-D-Arc Welderentals**

## **Calibration** Certificate

Unit Serial Number:LK220007GInstrument Number:61007IP Address:192.168.111.222

Date Tested:02-02-17Mac Address:00:0A:8D:00:EE:4F

Primary Standard								
	382 (F)		882 (F)		1382 (F)			
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display		
Channel 1	2.791	383	6.014	882	9.240	1382		
Channel 2	2.799	383	6.019	883	9.247	1382		
Channel 3	2.800	383	6.025	883	9.251	1383		
Channel 4	2.801	383	6.023	883	9.251	1383		
Channel 5	2.804	384	6.026	882	9.249	1383		
Channel 6	2.804	384	6.034	883	9.257	1384		

## Allowable difference in temp reading is +or- 5 Deg. F.

## **Instruments Used:**

Fluke 714 Thermocouple Calibrator SN: 1897018	Calibration Date: 11-04-16
Precision digital multimeter SN: US36049893	Calibration Date: 11-04-16
ProHeat 35 power supply used for signal output SN: MC100331G	Calibration date: 02-01-17

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.	X	Technician:	Jared Lamar
Calibration due date:	02-02-18		Signature