Red-D-Arc Welderentals

Calibration Certificate

Unit Serial Number: LK360096 Date Tested: 09-01-15

Instrument Number: 94572 Mac Address: 00:0A:8D:01:71:6C

IP Address: 192:168:111:222

Primary Standard							
	382 (F)		882 (F)		1382 (F)		
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display	
Channel 1	2.792	383	6.012	882	9.233	1382	
Channel 2	2.792	383	6.014	882	9.240	1382	
Channel 3	2.788	382	6.009	882	9.237	1382	
Channel 4	2.793	383	6.012	882	9.244	1383	
Channel 5	2.793	383	6.014	882	9.246	1383	
Channel 6	2.799	383	6.036	885	9.251	1384	

Allowable difference in temp reading is +or- 5 Deg. F.

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2969020 Calibration Date: 11-21-14

Precision digital multimeter SN: 30330025 Calibration Date: 01-23-15

ProHeat 35 power supply used for signal output SN: MC100327 Calibration date: 07-07-15

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.	x	Technician:	Jared Lamar
Calibration due date:	09-01-16		Signature