

Red-D-Arc Welderrentals

Calibration Certificate

Unit Serial Number: MA120025G
Instrument Number: 99291
IP Address: 192.168.111.222

Date Tested: FEB 21/17
Mac Address: 00:0A:8D:01:83:DB

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	2.789	382	6.015	883	9.237	1382
Channel 2	2.789	383	6.015	883	9.237	1382
Channel 3	2.786	382	6.015	882	9.239	1382
Channel 4	2.789	382	6.013	882	9.235	1382
Channel 5	2.786	382	6.015	882	9.239	1382
Channel 6	2.784	381	6.011	882	9.236	1382

Allowable difference in temp reading is +or- 5 Deg. F.

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2118055

Calibration Date: 9/21/16

Precision digital multimeter SN: 2142011

Calibration Date: 9/21/16

ProHeat 35 power supply used for signal output SN: ME270037G Calibration date: Feb 16/17

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Calibration due date:

Technician: D. Gurga
Signature