

# Red-D-Arc Welderrentals

## Calibration Certificate

Unit Serial Number: MC350424 G Date Tested: Jan 31 / 17.  
 Instrument Number: 8463721 Mac Address: 00:0A:8D:81:25:69  
 IP Address: 192.168.111.222

Primary Standard						
	382 (F) 194.5°c		882 (F) 473.5°c		1382 (F) 750.2°c	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	2.786	195°c	6.017	473.	9.239	751
Channel 2	2.784	194	6.032	474	9.229	750
Channel 3	2.786	194	6.019	473	9.231	749
Channel 4	2.787	195	6.018	473	9.232	750
Channel 5	2.788	194	6.013	473	9.235	749
Channel 6	2.782	195	6.032	474	9.235	749

Allowable difference in temp reading is +or- 5 Deg. F.

### Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2118055

Calibration Date: SEPT 21/16

Precision digital multimeter SN: 2142011

Calibration Date: SEPT 21/16

ProHeat 35 power supply used for signal output SN: MC200494

Calibration date: JAN 31/17

MC350424

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Calibration due date:

Technician: D. Burger

Signature