

Miller ProHeat 35

Unit Serial Number: MD170247G

Date Tested: 12-14-17

Certified Calibration

Primary Standard						
	382° (F) 2.787VDC@RC9		882° (F) 6.013VDC@RC9		1382° (F) 9.239VDC@RC9	
	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display
TC1	2.787	382	6.011	882	9.245	1381
TC2	2.793	383	6.017	882	9.241	1382
TC3	2.799	383	6.022	883	9.246	1382
TC4	2.797	383	6.015	883	9.249	1382
TC5	2.795	383	6.020	883	9.250	1383
TC6	2.794	382	6.015	882	9.246	1382
Allowable	2.749-2.825	+/- 3° F	5.975-6.051	+/- 3° F	9.201-9.277	+/- 3° F

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2969020

Calibration Date: 2-17-17

Precision Digital Voltage Meter SN: 3030025

Calibration Date: 2-17-17

Machine re-calibration due date: 12-14-18

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.



Technician: Brad Rowland

Date: 12-14-17

Signature